

Birch Lake 7-21-20



Eaton Analytical

LABORATORY REPORT

If you have any questions concerning this report, please do not hesitate to call us at (800) 332-4345 or (574) 233-4777.

This report may not be reproduced, except in full, without written approval from EEA.

STATE CERTIFICATION LIST

State	Certification	State	Certification
Alabama	40700	Missouri	880
Alaska	IN00035	Montana	CERT0026
Arizona	AZ0432	Nebraska	NE-OS-05-04
Arkansas	IN00035	Nevada	IN00035
California	2920	New Hampshire*	2124
Colorado	IN00035	New Jersey*	IN598
Colorado Radiochemistry	IN00035	New Mexico	IN00035
Connecticut	PH-0132	New York*	11398
Delaware	IN035	North Carolina	18700
Florida*	E87775	North Dakota	R-035
Georgia	929	Ohio	87775
Hawaii	IN035	Oklahoma	D9508
Idaho	IN00035	Oregon (Primary AB)*	4074
Illinois*	200001	Pennsylvania*	68-00466
Illinois Microbiology	17767	Puerto Rico	IN00035
Illinois Radiochemistry	IN00035	Rhode Island	LAO00343
Indiana Chemistry	C-71-01	South Carolina	95005
Indiana Microbiology	M-76-07	South Dakota	IN00035
Iowa	098	Tennessee	TN02973
Kansas*	E-10233	Texas*	T104704187-18-12
Kentucky	90056	Texas/TCEQ	TX207
Louisiana*	LA014	Utah*	IN00035
Maine	IN00035	Vermont	VT-8775
Maryland	209	Virginia*	460275
Massachusetts	M-IN035	Washington	C837
Michigan	9926	West Virginia	9927 C
Minnesota*	018-999-338	Wisconsin	999766900
Mississippi	IN035	Wyoming	IN035
EPA	IN00035		

*NELAP/TNI Recognized Accreditation Bodies



Eaton Analytical

110 South Hill Street
South Bend, IN 46617
Tel: (574) 233-4777
Fax: (574) 233-8207
1 800 332 4345

Laboratory Report

Client: USAF
Attn: Toru Kikunaga
18 AMDS/SGPN, Unit 5213
Bldg. 853
Kadena AFB, Okinawa 904-0203

Report: 494102
Priority: Standard Written
Status: Final
PWS ID: AK2372245
Alaska Lab ID # IN00035

Sample Information

EEA ID #	Client ID	Method	Collected Date / Time	Collected By:	Received Date / Time
4680022	GP20082	537.1	07/21/20 09:10	Client	07/23/20 09:30

Report Summary

Detailed quantitative results are presented on the following pages. The results presented relate only to the samples provided for analysis.

We appreciate the opportunity to provide you with this analysis. If you have any questions concerning this report, please do not hesitate to call Traci Chlebowski at (574) 233-4777.

Note: This report may not be reproduced, except in full, without written approval from EEA.

Traci Chlebowski - ASM

Authorized Signature

Title

08/21/2020

Date

Client Name: USAF

Report #: 494102

Client Name: USAF

Report #: 494102

Sampling Point: GP20082 Birch Lake well

0000 SPT 9

PWS ID: AK2372245

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
335-67-1	Perfluorooctanoic acid (PFOA) ✓	537.1	—	2.0	220	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
1763-23-1	Perfluorooctanesulfonic acid (PFOS) ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
375-73-5	Perfluorobutanesulfonic acid (PFBS) ✓	537.1	—	2.0	8.6	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
375-85-9	Perfluoroheptanoic acid (PFHpA) ✓	537.1	—	2.0	4.1	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
355-46-4	Perfluorohexanesulfonic acid (PFHxS) ✓	537.1	—	2.0	150	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
375-95-1	Perfluorononanoic acid (PFNA) ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
335-76-2	Perfluorodecanoic acid (PFDA) \$ ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
307-24-4	Perfluorohexanoic acid (PFHxA) \$ ✓	537.1	—	2.0	25	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
307-55-1	Perfluorododecanoic acid (PFDoA) \$ ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
72629-94-8	Perfluorotridecanoic acid (PFTrDA) \$ ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
2058-94-8	Perfluoroundecanoic acid (PFUnA) \$ ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
2991-50-6	N-ethyl Perfluorooctanesulfonamidoacetic acid \$ ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
2355-31-9	N-methyl Perfluorooctanesulfonamidoacetic acid \$ ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
13252-13-6	HFPO-DA/GenX \$ ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
919005-14-4	ADONA \$ ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
756426-58-1	9Cl-PF3ONS/F-53B Major \$ ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
763051-92-9	11Cl-PF3OUS/F-53B Minor \$ ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022
376-06-7	Perfluorotetradecanoic acid (PFTeDA) \$ ✓	537.1	—	2.0	< 2.0	ng/L	07/30/20 08:15	07/31/20 06:07	4680022

\$ The state of origin does not offer certification for this parameter.

† EEA has demonstrated it can achieve these report limits in reagent water, but can not document them in all sample matrices.

Reg Limit Type:	MCL	SMCL	AL
Symbol:	*	^	!

Lab Definitions

Continuing Calibration Check Standard (CCC) / Continuing Calibration Verification (CCV) / Initial Calibration Verification Standard (ICV) / Initial Performance Check (IPC) - is a standard containing one or more of the target analytes that is prepared from the same standards used to calibrate the instrument. This standard is used to verify the calibration curve at the beginning of each analytical sequence, and may also be analyzed throughout and at the end of the sequence. The concentration of continuing standards may be varied, when prescribed by the reference method, so that the range of the calibration curve is verified on a regular basis. CCL, CCM, and CCH are the CCC standards at low, mid, and high concentration levels, respectively.

Internal Standards (IS) - are pure compounds with properties similar to the analytes of interest, which are added to field samples or extracts, calibration standards, and quality control standards at a known concentration. They are used to measure the relative responses of the analytes of interest and surrogates in the sample, calibration standard or quality control standard.

Laboratory Duplicate (LD) - is a field sample aliquot taken from the same sample container in the laboratory and analyzed separately using identical procedures. Analysis of laboratory duplicates provides a measure of the precision of the laboratory procedures.

Laboratory Fortified Blank (LFB) / Laboratory Control Sample (LCS) - is an aliquot of reagent water to which known concentrations of the analytes of interest are added. The LFB is analyzed exactly the same as the field samples. LFBs are used to determine whether the method is in control. FBL, FBM, and FBH are the LFB samples at low, mid, and high concentration levels, respectively.

Laboratory Method Blank (LMB) / Laboratory Reagent Blank (LRB) - is a sample of reagent water included in the sample batch analyzed in the same way as the associated field samples. The LMB is used to determine if method analytes or other background contamination have been introduced during the preparation or analytical procedure. The LMB is analyzed exactly the same as the field samples.

Laboratory Trip Blank (LTB) / Field Reagent Blank (FRB) - is a sample of laboratory reagent water placed in a sample container in the laboratory and treated as a field sample, including storage, preservation, and all analytical procedures. The FRB/LTB container follows the collection bottles to and from the collection site, but the FRB/LTB is not opened at any time during the trip. The FRB/LTB is primarily a travel blank used to verify that the samples were not contaminated during shipment.

If applicable, the calculation of the matrix spike (MS) or matrix spike duplicate (MSD) percent recovery is as follows: $(MS \text{ or } MSD \text{ value} - \text{Sample value}) * 100 / \text{spike target} / \text{dilution factor} = \text{Recovery } \%$

Matrix Spike Duplicate Sample (MSD) / Laboratory Fortified Sample Matrix Duplicate (LFSMD) - is a sample aliquot taken from the same field sample source as the Matrix Spike Sample to which known quantities of the analytes of interest are added in the laboratory. The MSD is analyzed exactly the same as the field samples. Analysis of the MSD provides a measure of the precision of the laboratory procedures in a specific matrix. SDL, SDM, and SDH / LFSMDL, LFSMDM, and LFSMDH are the MSD or LFSMD at low, mid, and high concentration levels, respectively.

Matrix Spike Sample (MS) / Laboratory Fortified Sample Matrix (LFSM) - is a sample aliquot taken from field sample source to which known quantities of the analytes of interest are added in the laboratory. The MS is analyzed exactly the same as the field samples. The purpose is to demonstrate recovery of the analytes from a sample matrix to determine if the specific matrix contributes bias to the analytical results. MSL, MSM, and MSH / LFSML, LFSMM, and LFSMH are the MS or LFSM at low, mid, and high concentration levels, respectively.

Quality Control Standard (QCS) / Second Source Calibration Verification (SSCV) - is a solution containing known concentrations of the analytes of interest prepared from a source different from the source of the calibration standards. The solution is obtained from a second manufacturer or lot if the lot can be demonstrated by the manufacturer as prepared independently from other lots. The QCS sample is analyzed using the same procedures as field samples. The QCS is used as a check on the calibration standards used in the method on a routine basis.

Reporting Limit Check (RLC) / Initial Calibration Check Standard (ICCS) - is a procedural standard that is analyzed each day to evaluate instrument performance at or below the minimum reporting limit (MRL).

Surrogate Standard (SS) / Surrogate Analyte (SUR) - is a pure compound with properties similar to the analytes of interest, which is highly unlikely to be found in any field sample, that is added to the field samples, calibration standards, blanks and quality control standards before sample preparation. The SS is used to evaluate the efficiency of the sample preparation process.



Eaton Analytical

110 S. Hill Street
South Bend, IN 46617
T: 1.800.332.4345
F: 1.574.233.8207

Order # 401658
Batch # 492255
494102
8/11/2020

www.eurofins.com

CHAIN OF CUSTODY RECORD

Page 1 of 1

LAB NUMBER	DATE	TIME	AM	PM	COLLECTION	SAMPLER (Signature)	COMPLIANCE MONITORING		SAMPLING SITE	TEST NAME	POPULATION SERVED	STATE (SAMPLES ORIG)	PROJECT NAME	PRESERVATIVE CHECKS	CHLORINATED		# OF CONTAINERS	MATRIX CODE	TURNAROUND TIME
							Yes	No							YES	NO			
1	07/21/20	0910	X		GP20082 Birch Lake Well	Kathleen Leleiver		X	Groundwater	8193	Alaska	USAF-EC20-331				X	DW	SW	
2	07/21/20	1000	X		GP20083 FamCamp RV Park Well			X								X	DW	SW	
3	07/21/20	1030	X		GP20084 WTP POE										X		DW	SW	
4	07/21/20	1027	X		GP20085 WTP Field Blank												RW	SW	
5	07/21/20	1112	X		GP20066 Bldg 3408 Well A											X	DW	SW	
6	07/21/20	1140	X		GP20087 Bldg 3430 Well B											X	DW	SW	
7	07/21/20	1100	X		GP20088 Bldg 3225 Well E											X	DW	SW	
8	07/21/20	1125	X		GP20089 Bldg 3427 Well F											X	DW	SW	
9	07/21/20	1205	X		GP20090 Bldg 1301 Well 7											X	DW	SW	
10	07/21/20	1230	X		GP20091 Bldg 1346 Jet Engine Maintenance											X	DW	SW	
11	07/21/20	1230	X		BK20092 Bldg 1346 Trip Blank											X	RW	SW	

RELINQUISHED BY (Signature)	DATE	TIME	AM	PM	RECEIVED BY (Signature)	DATE	TIME	AM	PM	LAB COMMENTS
Kathleen Leleiver	7-22-20	0900								LAB RECEIVES THE POINT TO RETURN LOUSED PORTIONS OF EACH ADJUS SAMPLES TO CLIENT

RELINQUISHED BY (Signature)	DATE	TIME	AM	PM	RECEIVED FOR LABORATORY BY	DATE	TIME	AM	PM	CONDITIONS UPON RECEIPT (check one)
					Brenda Day	7/21/20	0930			<input checked="" type="checkbox"/> 100% <input type="checkbox"/> 75% <input type="checkbox"/> 50% <input type="checkbox"/> 25% <input type="checkbox"/> 10% <input type="checkbox"/> 0%

MATRIX CODES:
 DW - DRINKING WATER
 RW - REFRIGERANT WATER
 GW - GROUND WATER
 EW - EFFLUENT WATER
 SW - SURFACE WATER
 PW - POOL WATER
 WW - WASTE WATER

TURN-AROUND TIME (TAT) - SURCHARGES
 SW - Standard Within 15 working days) 0%
 RW - Rush Within 15 working days) 50%
 EW - Rush Within 15 working days) 75%
 SW - Standard Within 15 working days) 100%
 RW - Rush Within 15 working days) 125%
 EW - Rush Within 15 working days) 150%
 * Please call, expedited service not available for all testing